



Wash-Checks UH

Cleaning Monitor for Lumen Instruments

REF WC 110 (monitor)

WC 106 (Rigid Holder)
WC 107 (Flexible Holder)

FOR USE WITH ULTRASONIC WASHER

Directions for Use

Intended Use:

The Wash-Checks UH disposable cleaning monitor for lumened instruments is designed to monitor cleaning efficiency of Ultrasonic Cleaners.

Routine Use:

1. Open a Wash-Checks H or E holder and place a Wash-Checks UH monitor inside being careful not to touch the test soil. Close the holder by screwing the two sections together.
2. Attach the loaded holder to the connector using the luer lock in the Ultrasonic Washer or if no connector is available place the holder in the bottom of the basket.
3. Load the ultrasonic washer with a normal load of dirty hollow instruments.
4. Process per normal procedure.
5. Remove the Wash-Checks H or E holder, unscrew the two parts of the holder and remove the Wash-Checks UH monitor.
6. Examine Wash-Checks UH strip for visual cleanliness. If cleaning cycle was effective, BLUE SOIL should disappear.
7. Staple the Wash-Checks UH onto the Wash-Checks Record Sheet and fill in the information.



Wash-Checks UH

Cleaning Monitor for Lumen Instruments

REF WC 110 (monitor)

WC 106 (Rigid Holder)
WC 107 (Flexible Holder)

FOR USE WITH ULTRASONIC WASHER

Directions for Use

Intended Use:

The Wash-Checks UH disposable cleaning monitor for lumened instruments is designed to monitor cleaning efficiency of Ultrasonic cleaners.

Routine Use:

1. Open a Wash-Checks H or E holder and place a Wash-Checks UH monitor inside being careful not to touch the test soil. Close the holder by screwing the two sections together.
2. Attach the loaded holder to the connector using the luer lock in the Ultrasonic Washer or if no connector is available place the holder in the bottom of the basket.
3. Load the ultrasonic washer with a normal load of dirty hollow instruments.
4. Process per normal procedure.
5. Remove the Wash-Checks H or E holder, unscrew the two parts of the holder and remove the Wash-Checks UH monitor.
6. Examine Wash-Checks UH strip for visual cleanliness. If cleaning cycle was effective, BLUE SOIL should disappear.
7. Staple the Wash-Checks UH onto the Wash-Checks Record Sheet and fill in the information.



Wash-Checks UH

Cleaning Monitor for Lumen Instruments

REF WC 110 (monitor)

WC 106 (Rigid Holder)
WC 107 (Flexible Holder)

FOR USE WITH ULTRASONIC WASHER

Directions for Use

Intended Use:

The Wash-Checks UH disposable cleaning monitor for lumened instruments is designed to monitor cleaning efficiency in ultrasonic cleaners.

Routine Use:

1. Open a Wash-Checks H or E holder and place a Wash-Checks UH monitor inside being careful not to touch the test soil. Close the holder by screwing the two sections together.
2. Attach the loaded holder to the connector using the luer lock in the Ultrasonic Washer or if no connector is available place the holder in the bottom of the basket.
3. Load the ultrasonic washer with a normal load of dirty hollow instruments.
4. Process per normal procedure.
5. Remove the Wash-Checks H or E holder, unscrew the two parts of the holder and remove the Wash-Checks H monitor.
6. Examine Wash-Checks UH strip for visual cleanliness. If cleaning cycle was effective, BLUE SOIL should disappear.
7. Staple the Wash-Checks UH onto the Wash-Checks Record Sheet and fill in the information.



Wash-Checks UH

Cleaning Monitor for Lumen Instruments

REF WC 110 (monitor)

WC 106 (Rigid Holder)
WC 107 (Flexible Holder)

FOR USE WITH ULTRASONIC WASHER

Directions for Use

Intended Use:

The Wash-Checks UH disposable cleaning monitor for lumened instruments is designed to monitor cleaning efficiency in ultrasonic cleaners.

Routine Use:

1. Open a Wash-Checks H or E holder and place a Wash-Checks UH monitor inside being careful not to touch the test soil. Close the holder by screwing the two sections together.
2. Attach the loaded holder to the connector using the luer lock in the Ultrasonic Washer or if no connector is available place the holder in the bottom of the basket.
3. Load the ultrasonic washer with a normal load of dirty hollow instruments.
4. Process per normal procedure.
5. Remove the Wash-Checks H or E holder, unscrew the two parts of the holder and remove the Wash-Checks H monitor.
6. Examine Wash-Checks UH strip for visual cleanliness. If cleaning cycle was effective, BLUE SOIL should disappear.
7. Staple the Wash-Checks UH onto the Wash-Checks Record Sheet and fill in the information.



Interpretation of Results:

Examine the aluminum strip for presence of any test soil remaining. The blue soil should disappear in an efficient ultrasonic wash cycle. The presence of a large amount of soil remaining on the strip means a serious deficiency in ultrasonic washer efficiency. A small amount of soil remaining indicates that conditions are slightly deficient (marginal).

Storage and Handling:

Store Wash-Checks UH in the tube at room temperature. Keep away from disinfectants and fumes. Remove only the strips needed for the test at hand and re-seal the vial. The expiry date is printed on the product packaging.

Suggested Record Keeping:

If you wish to retain the Wash-Checks UH Monitor as part of a permanent ultrasonic cycle record, the following record keeping method is suggested:



After the cycle is completed, collect the Wash-Checks UH strips and staple each strip to the Wash-Checks Cleaning Record form (MK-105) for archiving. On the form, record the Date, Initials, the time or cycle number, and check the appropriate Wash-Checks results (pass/fail/marginal) box.

Interpretation of Results:

Examine the aluminum strip for presence of any test soil remaining. The blue soil should disappear in an efficient ultrasonic wash cycle. The presence of a large amount of soil remaining on the strip means a serious deficiency in ultrasonic washer efficiency. A small amount of soil remaining indicates that conditions are slightly deficient (marginal).

Storage and Handling:

Store Wash-Checks UH in the box at room temperature. Keep away from disinfectants and fumes. Remove only the strips needed for the test at hand and re-seal the vial. The expiry date is printed on the product packaging.

Suggested Record Keeping:

If you wish to retain the Wash-Checks UH Monitor as part of a permanent ultrasonic cycle record, the following record keeping method is suggested:



After the cycle is completed, collect the Wash-Checks UH strips and staple each strip to the Wash-Checks Cleaning Record form (MK-105) for archiving. On the form, record the Date, Initials, the time or cycle number, and check the appropriate Wash-Checks results (pass/fail/marginal) box.

Interpretation of Results:

Examine the aluminum strip for presence of any test soil remaining. The blue soil should disappear in an efficient ultrasonic wash cycle. The presence of a large amount of soil remaining on the strip means a serious deficiency in ultrasonic washer efficiency. A small amount of soil remaining indicates that conditions are slightly deficient (marginal).

Storage and Handling:

Store Wash-Checks UH in the box at room temperature. Keep away from disinfectants and fumes. Remove only the strips needed for the test at hand and re-seal the vial. The expiry date is printed on the product packaging.

Suggested Record Keeping:

If you wish to retain the Wash-Checks UH Monitor as part of a permanent wash cycle record, the following record keeping method is suggested:



After the cycle is completed, collect the Wash-Checks UH strips and staple each strip to the Wash-Checks Cleaning Record form for archiving. On the form, record the Date, Initials, the time or cycle number, and check the appropriate Wash-Checks results (pass/fail/marginal) box.

Interpretation of Results:

Examine the aluminum strip for presence of any test soil remaining. The blue soil should disappear in an efficient ultrasonic wash cycle. The presence of a large amount of soil remaining on the strip means a serious deficiency in ultrasonic washer efficiency. A small amount of soil remaining indicates that conditions are slightly deficient (marginal).

Storage and Handling:

Store Wash-Checks UH in the box at room temperature. Keep away from disinfectants and fumes. Remove only the strips needed for the test at hand and re-seal the vial. The expiry date is printed on the product packaging.

Suggested Record Keeping:

If you wish to retain the Wash-Checks UH Monitor as part of a permanent wash cycle record, the following record keeping method is suggested:



After the cycle is completed, collect the Wash-Checks UH strips and staple each strip to the Wash-Checks Cleaning Record form for archiving. On the form, record the Date, Initials, the time or cycle number, and check the appropriate Wash-Checks results (pass/fail/marginal) box.