


Disposable Steam Biological Test Pack


Catalog No. BI 103

UNEXPOSED

Color of "PASS" will change from purple to green when proper sterilization conditions have been met.



PASS



INSUFFICIENT

