


# Disposable Steam Biological Test Pack


Catalog No. BI 103/INT



## UNEXPOSED

Color of "PASS" will change from purple to green when proper sterilization conditions have been met.

PASS



## INSUFFICIENT